



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

### **Scale South**

**313 Commerce Drive, Martinez, GA 30907  
120-A Pidgeon Bay Road, Summerville, SC 29483  
218 Bourne Boulevard Suite G, Savannah, GA 31408**

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

### **ISO/IEC 17025:2017**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

### **Calibration of Weighing Devices** *(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President

*Initial Accreditation Date:*

July 15, 2003

*Issue Date:*

March 25, 2021

*Expiration Date:*

June 30, 2023

*Accreditation No.:*

59247

*Certificate No.:*

L21-190

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjllabs.com](http://www.pjllabs.com)*



# Certificate of Accreditation: Supplement

## Scale South

313 Commerce Drive, Martinez, GA 30907  
 120-A Pidgeon Bay Road, Summerville, SC 29483  
 218 Bourne Boulevard Suite G, Savannah, GA 31408  
 Contact: Stacy Moore Phone: 706-855-1111

Accreditation is granted to the facility to perform the following calibrations:

### Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Analytical Balances <sup>FO</sup>	1 mg to 320 g (Res. = 0.1 mg)	$(2 \times 10^{-4} + 2.7 \times 10^{-8} \text{Wt}) \text{ g}$	Class 1 Weights ASTM 898:2020 Quality Procedure 1.D
Top Load Balances <sup>FO</sup>	0.05 g to 20 kg (Res. = 0.01 g)	$(1.16 \times 10^{-2} + 1.15 \times 10^{-5} \text{Wt}) \text{ kg}$	Class F Weights NIST Handbook 44 Quality Procedures 1.A through 1.G
Bench Scales <sup>FO</sup>	0.05 kg to 30 kg (Res. = 0.01 kg)	$(1.16 \times 10^{-2} + 1.7 \times 10^{-5} \text{Wt}) \text{ kg}$	
	0.2 kg to 150 kg (Res. = 0.05 kg)	$(5.78 \times 10^{-2} + 1.7 \times 10^{-5} \text{Wt}) \text{ kg}$	
Floor Scales <sup>FO</sup>	1 kg to 907.185 kg (Res. = 0.2 kg)	$(2.31 \times 10^{-1} + 1.06 \times 10^{-4} \text{Wt}) \text{ kg}$	
	(2 lb to 2 000 lb) (Res. = 0.5 lb)	$(5.77 \times 10^{-1} + 2.23 \times 10^{-5} \text{Wt}) \text{ lb}$	
	2 kg to 2 267.962 kg (Res. = 0.5 kg)	$(5.77 \times 10^{-1} + 2.5 \times 10^{-5} \text{Wt}) \text{ kg}$	
	(2 lb to 5 000 lb) (Res. = 0.5 lb)	$(5.77 \times 10^{-1} + 4.78 \times 10^{-5} \text{Wt}) \text{ lb}$	
	4 kg to 4 535.924 kg (Res. = 1 kg)	$(1.154 7 + 2.5 \times 10^{-5} \text{Wt}) \text{ kg}$	
	8 lb to 10 000 lb (Res. = 2 lb)	$(2.309 2 + 2.73 \times 10^{-5} \text{Wt}) \text{ lb}$	
	8 kg to 9 071.847 kg (Res. = 2 kg)	$(2.309 3 + 2.51 \times 10^{-5} \text{Wt}) \text{ kg}$	
	20 lb to 20 000 lb (Res. = 5 lb)	$(5.773 1 + 2.23 \times 10^{-5} \text{Wt}) \text{ lb}$	
Tank and Hopper Scales <sup>FO</sup>	20 kg to 27 215 kg (Res. = 5 kg)	$(5.773 6 + 2.94 \times 10^{-5} \text{Wt}) \text{ kg}$	
	40 lb to 60 000 lb (Res. = 10 lb)	$(11.545 8 + 3.2 \times 10^{-5} \text{Wt}) \text{ lb}$	
	80 kg to 158 757.3 kg (Res. = 20 kg)	$(23.082 2 + 1.48 \times 10^{-4} \text{Wt}) \text{ kg}$	
	200 lb to 350 000 lb (Res. = 50 lb)	$(57.707 4 + 1.38 \times 10^{-4} \text{Wt}) \text{ lb}$	
Truck Scales <sup>FO</sup>	40 kg to 90 718.474 kg (Res. = 10 kg)	$(11.545 3 + 4.46 \times 10^{-5} \text{Wt}) \text{ kg}$	
	80 lb to 200 000 lb (Res. = 20 lb)	$(23.090 2 + 4.78 \times 10^{-5} \text{Wt}) \text{ lb}$	



# Certificate of Accreditation: Supplement

## Scale South

313 Commerce Drive, Martinez, GA 30907  
120-A Pidgeon Bay Road, Summerville, SC 29483  
218 Bourne Boulevard Suite G, Savannah, GA 31408  
Contact: Stacy Moore Phone: 706-855-1111

Accreditation is granted to the facility to perform the following calibrations:

### Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Wheel Load Weigher <sup>FO</sup>	80 kg to 9 080 kg (Res. = 20 kg)	24 kg	Class F Weights NIST Handbook 44 Quality Procedures 1.A through 1.G
	200 lb to 20 000 lb (Res. = 50 lb)	57.781 3 lb	
Rail Scales <sup>O</sup>	80 kg to 181 436.948 kg (Res. = 20 kg)	(23.090 5 + 4.46 x 10 <sup>-5</sup> Wt) kg	
	200 lb to 400 000 lb (Res. = 50 lb)	(57.727 + 4.05 x 10 <sup>-5</sup> Wt) lb	
Crane Scales <sup>FO</sup>	20 kg to 27 215 kg (Res. = 5 kg)	(5.777 36 + 2.94 x 10 <sup>-5</sup> Wt) kg	
	40 lb to 60 000 lb (Res. = 10 lb)	(11.545 8 + 3.2 x 10 <sup>-5</sup> Wt) lb	

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor *k* (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration performed by the laboratory with the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations. Example: Outside Micrometer<sup>O</sup> would mean that the laboratory performs this calibration onsite at the customer's location.
4. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer<sup>FO</sup> would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.



## *Certificate of Accreditation: Supplement*

### **Scale South**

313 Commerce Drive, Martinez, GA 30907  
120-A Pidgeon Bay Road, Summerville, SC 29483  
218 Bourne Boulevard Suite G, Savannah, GA 31408  
Contact: Stacy Moore Phone: 706-855-1111

*Accreditation is granted to the facility to perform the following calibrations:*

5. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
6. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.

